

Artron Certificate of Analysis

Product Name	COVID-19 IgM/IgG Antibody Test □ Strip ☑ Cassette □ Uncut Sheet			Catalog No.		A03-51-322			
Lot No.	N19-500320		Qua	antity	25		Shelf L	ife	2021-08
Store	2-30°C		Release Date 2020		0-03-24				
Appearance	☐ Strip 3mm Single test, Sealed in desiccated aluminum pouch								
	☑ Cassette	Single test, Sealed in desiccated aluminum pouch							
	☐ Uncut Sheet	Sealed in desiccated aluminum pouch							

Applicable Inspection and Evaluation of specification:

Physical Inspection:

Test		D 1.		
	☐ Strip	☑ Cassette	☐ Uncut Sheet	Results
Width	3.0±0.1mm	/	/	Acceptable
Appearance	AQL=0.65	AQL=0.65	/	Acceptable
Label	AQL=0.65	AQL=0.65	AQL=0.65	Acceptable
Pouch	AQL=0.65	AQL=0.65	AQL=0.65	Acceptable
Desiccant in pouches	AQL=0.65	AQL=0.65	AQL=0.65	Acceptable

Note: AQL means Acceptance Quality Limit, quality level that is the worst tolerable process average when a continuing series of lots is submitted for acceptance sampling.

Performance Characteristic Inspection:

	Test	Specification	Results
Sensitivity	COVID-19 P1 (IgM Positive)	Positive	
	COVID-19 P2 (IgG Positive)	Positive	Pass
	COVID-19 P3 (IgM/IgG Positive)	Positive	
Specificity	Negative enhancement sample	Negative	Daga
	COVID-19 Negative control, N1-N8	All negative	Pass

Note: All data are from QC testing unless otherwise indicated.

Safety Certification:

- 1. Appropriate inspection has been applied to this lot, and to our knowledge, specific defects do not exceed the maximum allowable quantities as listed in the applicable specifications.
- 2. This lot has been manufactured and packaged in accordance with the relevant regulations for diagnostic and medical devices industry in Canada.
- There are no biohazard materials utilized in this product. No radioactive materials, no narcotic reagents, neither their derivatives, no virus neither their reagent nor metabolic by-products.

Approved by: **Date:** <u>2020-03-24</u>